



Procurement of ORTHOPEDIC FRACTURE TABLE  
IB No. 2020-09-68R

This Addendum No. 01 is issued as an amendment to the Bidding Documents for the *Procurement of ORTHOPEDIC FRACTURE TABLE*. This shall form as an integral part of the said documents which shall be submitted during the Opening of Bids on October 26, 2020.

1. Technical Specifications:

ITEMS :

- A. Lateral support set WITH clamp for attachment to table (minimum of three components; back support, buttock support and sternum support), adjustable and padded

Is hereby amended and shall be read as :

Lateral support set or **its equivalent** WITH clamp for attachment to table (minimum of three components; back support, buttock support and sternum support), adjustable and padded

- B. Extension leg plate

Is hereby amended and shall be read as :

Extension leg plate or **its equivalent**

ATTACHMENTS :

- C. Spine frame padded, head support, and four point abdominal support allows supine, lateral and prone position; headrest allows airway access and visibility; radiolucent material for unrestricted imaging and weight capacity at least 200kg

Is hereby amended and shall be read as :

Spine frame or equivalent name, padded, head support, and four point abdominal support allows supine, lateral and prone position; headrest allows airway access and visibility; radiolucent material for unrestricted imaging and weight capacity at least 200 kg

2. Other : Terms and Conditions

Please refer to the attached amended : Terms and Conditions

*[Signature]* 10/14/2020

For your information and guidance.

*[Signature]*  
**LINO SANTIAGO PABILLO, MD, FPCR, MMHA, PHD**  
Chairman, Bids and Awards Committee

## Terms and Condition

1. Shall submit a copy of the Certificate of Exclusive Distributorship or Authority to Distribute from the Manufacturer/ Exclusive Distributor of the specific item offered to QMMC. Original certificate must be present during post-qualification.
2. Shall submit a valid and updated Certificate of Conformance for the specific equipment being offered issued by the following standards: Original certificate must be present during post-qualification.
  - a. ISO13485:2016
  - b. IEC 60601-1 or its equivalent in Medical Electrical Equipment Safety standards
3. Shall submit valid and updated Certificate of Good Performance for Medical Equipment
  - a. For bidders with 2019 contracts with QMMC, certificates of Good Performance issued by the procurement office and signed by the HOPE of QMMC is required
  - b. For new suppliers or without 2019 contract with QMMC, certificate of good performance shall be submitted from at least one (1) government agency
4. Shall submit bidder's Company Profile showing its organizational structure, particularly indicating its after sales division/section including the names, contact numbers, and experience of at least two (2) trained technical personnel for the equipment being offered presently employed in the company. Training Certificates shall be included in the bidding documents.
5. Shall submit Certificate from Manufacturer guaranteeing availability of parts for the next two (2) years after the warranty period, and guaranteeing delivery of equipment and all Accessories as stipulated in Section VI Schedule of Requirements.
6. Shall submit brochure/s of the equipment according to exact model offered that will support the specification declared under section VII and showcases the main technical requirements of the end user.
7. Shall submit certificate from manufacturer that all terms and conditions stated in the contract shall be honored by the manufacturer in the event that a change of exclusive distributorship would occur during the duration of the contract.
8. Shall submit a duly notarized DECLARATION of the following: (items A to F)
  - a. The bidder shall submit Manual in English Language upon delivery:
    - Operational Manual/ Service manual: 1 copy per unit
  - b. The bidder shall provide warranty as follows:  
(warranty shall commence after commissioning and acceptance by the end-user.)
    - 1) Service/Maintenance- three (3) years
    - 2) Parts and Accessories- two (2) years

A handwritten signature in blue ink, appearing to read "Agustin M. Agustin", is written over a horizontal line.



c. The bidder has certified Service Engineer/Specialist, on call, 24/7. Bidder shall perform preventive maintenance quarterly whichever comes first, and provide certificate of preventive maintenance. A copy of the engineer/technicians certificate of training should be attached to the bidding documents.

d. The bidder shall provide orientation and training with certificate/s for end user and biomedical personnel on equipment operation and preventive maintenance.

e. That any packaging material made of styropore or other non-biodegradable materials shall be collected by the winning bidder after installation.

f. The bidder shall conduct product demo within the hospital or arrange for a site visit during post-qualification. Demo unit should be available or site visit should be arranged within 5-calendar days upon receipt of notification and shall be used for a maximum of seven (7) calendar days (if with demo unit). Demo unit should be identical with the offered product declared technical specification under Section VII.

\*During demo/ on site visit. At least the following should be present and of the same specifications;

a. Electro-hydraulic operating room table

b. Orthopedic Traction Set [carbon fiber, table adapters, two telescopic bars/traction system by mechanical lever, traction boots with up/down/left/right movement and rotation capability around its own axis, padded pubic/perineal/sacral support] \*with storage cart

c. Arthroscopy Shoulder Support [At least 5-stage back rest adjustment for beach chair position capability, adjustable head holder/head fixing section, head/chin straps, ball-joint adjustment for height and rotation, 3 section back plate/back contour for shoulder accessibility, removable side sections for shoulder accessibility; side sections at least 15" in length; padded body fixing strap]; weight support at least 200kg. \*with shoulder support cart

d. Arm support with arm/wrist strap, padded; ball joint; at least 20" in length and 5" in width WITH clamp for attachment to table

e. Stirrups/knee crutches, padded; with straps WITH clamp for attachment to table

f. Foot support/board WITH clamp for attachment to table, at least 10" in both length and height

g. Lateral support set or its equivalent WITH clamp for attachment to table (minimum of three components; back support, buttock support and sternum support), adjustable and padded

9. Submit an electronic copy (flash drive/USB) of the bidders' Technical Proposals in Microsoft Excel or Word, and a soft copy of product brochures in PDF or JPEG file.

