



BID BULLETIN NO. 1
JULY 12, 2021

PROCUREMENT OF UVC STERILIZER LAMP HEAVY DUTY (REBID)
IB NO. 2021-06-27R

This Bid Bulletin is being issued to clarify the issue and questions raised during the Pre-Bidding Conference held June 30, 2021; 2:00pm at Boardroom, Quirino Memorial Medical Center, Project 4, Quezon City. This Bid Bulletin will form an integral part of the bidding document for the Procurement of UVC Sterilizer Lamp Heavy Duty (Rebid) under IB 2021-06-027R. The following are the changes in the bidding documents:

SECTION VII. TECHNICAL SPECIFICATIONS

PARTICULAR	CHANGES
Input Current: Max. 15Amp.	As manufacturer's specification
Power: Max of 1800Watt	Power: at least 120Watt
Lamp lifetime: at least 12,000hrs	Lamp lifetime: at least 2000 per lamp
To provide the ff per unit: <ul style="list-style-type: none">• 1-goggles with UV protection• additional 4 Ozone-free lamps back up• 70 meters reel extension cord	To provide the ff per unit: <ul style="list-style-type: none">• 3 pcs. goggles with UV protection• additional 4 Ozone-free lamps back up• 3 units of 15 meters reel extension cord heavy duty

The deadline for Submission and Opening of Bids is scheduled on 12 July 2021; 2:00pm at Boardroom, Quirino Memorial Medical Center, Project 4, Quezon City.

All other provisions of bidding documents which are not affected shall remain in effect.

For your guidance and information of all concerned.

LEONARDO E. CONCEPCION, JR., REE, MAPA, PhD
Chairman, Bids and Awards Committee