

Republic of the Philippines
Department of Health
QUIRINO MEMORIAL MEDICAL CENTER
Project 4, Quezon City

**SUPPLEMENTAL / BID BULLETIN
NO. 01**

**Procurement of Picture Archiving and Communication System (PACS)
IB no. 2019-05-24**

This Addendum No. 01 is issued as an amendment to the Bidding Documents for the procurement of **Procurement of Picture Archiving and Communication System (PACS)**. This shall form as an integral part of the said documents which shall be submitted during the Opening of Bids on **June 11, 2019**.

1. Section VI. Schedule of Requirements.

- Delivered, Weeks/Months: 15 days upon receipt of Notice to Proceed (for Hardwares). *is hereby amended and shall be read as: 30 calendar days* upon receipt of Notice to Proceed (for Hardwares).

2. Section VII. Technical Specification (Annex A)

QMMC TECHNICAL SPECIFICATION:

- UPS, 3 KVA, 220V, 60Hz *is hereby added*.

TERMS OF REFERENCE:

- 16. PACS-RIS system must be capable of off-site remote access by end user *is hereby amended and shall be read as*; PACS-RIS system must be capable of off-site remote access by end user via application and latest version of web browser.


OTHER TERMS AND CONDITIONS:

Shall submit a duly notarized Declaration of the following:


1. Shall provide 3 years warranty and quarterly preventive maintenance of the PACS system including parts, system update and upgrade
 - a. Upon expiration of 3 years warranty the cost of preventive maintenance services including system upgrade and update shall not exceed 20% contract price.
2. Shall provide free access for the RIS, either application-based or web-based access, for 3 years including system software and application update and upgrade.
3. Customization period covers from date of the installation up to three (3) months
4. Software modification period covers from system full implementation to nine (9) months.
5. PACS-RIS complaints should be remotely addressed by service engineer/technician/application specialist within 1 hour. If the problem cannot be remotely fixed within 1 hour, the service engineer/technician/application specialist should be physically present within the hospital premises within 3 hours.
6. Should provide training for end-users (radiation technologists, residents, consultants, and designated PACS-RIS administrator) and personnel of the information system of the hospital for at least 1 month, or more as to be requested by the end-user. Server

administration/virtualization, basic troubleshooting, and all IT related skills which the supplier used to implement the application must be echoed to specified ICMS staff and PACS-RIS administrator.

7. All changes/enhancements/updates in the system or in the equipment must be communicated with the hospital ICMS.
8. Automated daily backup of the entire PACS-RIS database, including all patient information and complete text of results (both initial and finalized), with on-site and other backup systems.
9. Full payment shall be made after full implementation (including customization) and Acceptance of the PACS-RIS.
10. Lowest bidder must be willing to conduct site visit of their existing installation
11. Winning bidder must shoulder all fees for the data migration from the existing PACS

 ➤ Other parameter not stated remains the same

For your information and guidance.


ISABEL ELONA, MSc
Chairman, Bids and Awards Committee