



SUPPLEMENTAL/BID BULLETIN
NO. 01
Procurement of Echocardiography Machine
IB#2019-07-33


This Addendum No. 01 is issued in addition to the Bidding Documents. This shall form an integral part of the Technical Documents to be submitted in the Opening of Bids.

Upon deliberation of the BAC which was conducted last August 08, 2019, 11:00am the following issues were resolved:

1. Annex A

- iScan control for 2D/Doppler/color doppler automatic optimization *is hereby amended and shall be read as:* Scan control or one button control for 2D/Doppler/color doppler automatic optimization
- Internal lithium battery with approximately 1 hour battery life under continuous use *is hereby amended and shall be read as:* Internal lithium battery with approximately 45-1 hour battery life under continuous use
- USB hub with cables to B/W printer, color printer, USB *is hereby amended and shall be read as:* USB hub with cables to B/W printer or color printer, USB
- 500GB hard drive memory storage internal slot CD/DVD/RW Drive *is hereby amended and shall be read as:* 500GB hard drive or SSD memory storage internal slot CD/DVD/RW Drive
- Probes: *is hereby amended and shall be read as:* Probes: manufacturer's operating frequency
- Black and white thermal printer, 110mmx20m *is hereby added.*

For your information and guidance.


OFELIA R. DE LEON, MD, FPPS, FPNSE, MMHA
Chairperson, Bids and Awards Committee (Equipment)