



SUPPLEMENTAL/BID BULLETIN

NO. 01

**Procurement of Cone Beam Computed Tomography (Digital Dental/Panoramic & Cephalometric)
Rebid
IB#2019-07-34**

This Addendum No. 01 is issued in addition to the Bidding Documents. This shall form an integral part of the Technical Documents to be submitted in the Opening of Bids.

Upon deliberation of the BAC which was conducted last August 08, 2019, 1:00 pm the following issues were resolved:


1. Annex A


- 2D Images
 - Pixel Size : 48 μ m *is hereby amended and shall be read as:* Pixel Size : 70-90 μ m
- Computer Workstation
 - at least 23.6 inches full HD Medical Grade 3 megapixel Monitor *is hereby amended and shall be read as:* at least 23.6 inches full HD Medical Grade 2-3 megapixel Monitor

2. Terms and Conditions:

- I. Shall provide preventive maintenance and calibration quarterly or if needed whichever comes first and shall provide certificate of calibration *is hereby amended and shall be read as:* Shall provide preventive maintenance and calibration quarterly or if needed whichever comes first and shall provide certificate and sticker of calibration.

For your information and guidance.


OFELIA R. DE LEON, MD, FPPS, FPNSP, MMHA
Chairperson, Bids and Awards Committee (Equipment)


DRS CB-07-19