



**SUPPLEMENTAL/BID BULLETIN
NO. 02**

**Procurement of Cone Beam Computed Tomography (Digital Dental/Panoramic & Cephalometric)
Rebid
IB#2019-07-34**


This Addendum No. 02 is issued in addition to the Bidding Documents. This shall form an integral part of the Technical Documents to be submitted in the Opening of Bids.

Upon deliberation of the BAC which was conducted last August 08, 2019, 1:00 pm the following issues were resolved:

1. Annex A

- 2D Images
 - Pixel Size : 70-90 μm *is hereby amended and shall be read as:* Pixel Size : 70-100 μm

For your information and guidance.


OFELIA R. DE LEON, MD, FPPS, FPNSP, MMHA
Chairperson, Bids and Awards Committee (Equipment)


MEB
60-64505 DOR 8/14/19
ORS