



**SUPPLEMENTAL/BID BULLETIN  
NO. 01  
Procurement of Laryngoscope Set (Rebid)  
IB#2019-07-35**

This Addendum No. 01 is issued in addition to the Bidding Documents. This shall form an integral part of the Technical Documents to be submitted in the Opening of Bids.

Upon deliberation of the BAC which was conducted last August 08, 2019, 2:00 pm the following issues were resolved:

1. Bid Data Sheet

- 5.4. For this purpose, similar contracts shall refer Laryngoscope Set *is hereby amended and shall be read as:* For this purpose, similar contracts shall refer ENT Instruments and Equipment


2. Annex A

- A) Biopsy forceps, flexible, double action jaws, oval diameter 1mm, length 60cm *is hereby amended and shall be read as:* Biopsy forceps, flexible, double action jaws, oval diameter 1-2mm, length 60cm
- B) Grasping forceps, flexible, double action jaws, diameter 1.7mm, length 60cm *is hereby amended and shall be read as:* Grasping forceps, flexible, double action jaws, diameter 1-2mm, length 60cm
- Tele-Laryngoscope, with integrated lateral telescope 70°, angle of view 50°, 5.8mm, length 20cm-23cm *is hereby amended and shall be read as:* Tele-Laryngoscope, with integrated lateral telescope 70°, angle of view 50°, 5.8mm, length 19cm-23cm
- Microlaryngeal Hook 45 degrees, length 23-25cm *is hereby amended and shall be read as:* Microlaryngeal hook or kleinsasser hook sharp, length 23-25cm

3. Terms and Conditions:

- j. That the bidder is willing to conduct product demo within the hospital during post-qualification. The bidder must have proof of calibration for its demo unit. The demo unit's specifications should reflect the bidder's offer *is hereby amended and shall be read as:* That the bidder is willing to conduct product demo for the equipment within the hospital during post-qualification. The bidder must have proof of calibration for its demo unit. The demo unit's specifications should reflect the bidder's offer. For instruments, bidder may present brochures for evaluation during post qualification.

For your information and guidance.

  
**OFELIA R. DE LEON, MD, FPPS, FPNP, MMHA**  
Chairperson, Bids and Awards Committee (Equipment) 