

Republic of the Philippines  
Department of Health  
**QUIRINO MEMORIAL MEDICAL CENTER**  
Project 4, Quezon City

---

SUPPLEMENTAL / BID BULLETIN

NO. 01

**Procurement of Electronic Pneumatic Tourniquet Machine and Cuffs for Pedia and Adult (Re-Bid)**  
**IB no. 2019-09-41R**

This Addendum No. 01 is issued as an amendment to the Bidding Documents for the procurement of **Electronic Pneumatic Tourniquet Machine and Cuffs for Pedia and Adult (Re-Bid)**. This shall form as an integral part of the said documents which shall be submitted during the Opening of Bids on **October 21, 2019**.

1. Bid Data Sheet:

- ITB Clause 1.2: The name of the Contract is *Procurement of Electronic Pneumatic Tourniquet Machine dual Port and Cuffs for Pedia and Adult (Re-Bid)* is hereby amended and shall be read as: The name of the Contract is *Procurement of Electronic Pneumatic Tourniquet Machine and Cuffs for Pedia and Adult (Re-Bid)*
- ITB Clause 2. The Government of the Philippines (GOP) through *its HFEP / INCOME FUND* in the amount of *EIGHT MILLION FOUR HUNDRED EIGHTY PESOS ONLY(Php8,480,000.00)* is hereby amended and shall be read as: The Government of the Philippines (GOP) through *its HFEP FUND* in the amount of *NINE HUNDRED EIGHTY THOUSAND PESOS ONLY(Php980,000.00)*

2. Annex A:

Items:

- Electronic pneumatic tourniquet machine system unit *is hereby amended and shall be read as:* Dual Port Electronic pneumatic tourniquet machine system unit
- Tourniquet cuff 30-39 inches with ports *is hereby amended and shall be read as:* Reusable Tourniquet cuff 30-39 inches with ports, 2 pcs.
- Tourniquet cuff 12-30 inches with ports *is hereby amended and shall be read as:* Reusable Tourniquet cuff 12-30 inches with ports, 2 pcs.
- Tourniquet cuff 9-12 inches with ports *is hereby amended and shall be read as:* Reusable Tourniquet cuff 9-12 inches with ports, 2 pcs.

Specification:

- Cuff pressure range (50-600mmHg), adjustable with minimum 1mmHg increments *is hereby amended and shall be read as:* Cuff pressure range (50-400mmHg), adjustable with minimum 1 mmHg increments
- Pressure accuracy +/- 3mmHg *is hereby amended and shall be read as:* Pressure accuracy +/- 3-4mmHg
- Time Alert Set Ranges: 0-300 minutes, in 1 minute increments *is hereby amended and shall be read as:* Time Alert Set Ranges: 0-200 minutes, in 1 minute increments

Republic of the Philippines  
Department of Health  
**QUIRINO MEMORIAL MEDICAL CENTER**  
Project 4, Quezon City

---

Other information not stated remains the same.

For your information and guidance.

  
Ofelia R. De Leon, MD  
Chairman, Bids and Awards Committee (Equipment)

