

Republic of the Philippines  
Department of Health  
**QUIRINO MEMORIAL MEDICAL CENTER**  
Project 4, Quezon City

---

**SUPPLEMENTAL / BID BULLETIN**  
NO. 01

**Procurement of Supply and Installation of Oxygen Flow meter, Air flow meter and Vacuum Regulator (1 Lot)**  
**IB no. 2019-10-47**

This Addendum No. 01 is issued as an amendment to the Bidding Documents for the procurement of **Procurement of Supply and Installation of Oxygen Flow meter, Air flow meter and Vacuum Regulator (1 Lot)**. This shall form as an integral part of the said documents which shall be submitted during the Opening of Bids on **November 6, 2019**.

1. Section III. Bid Data Sheet:

- ITB Clause 5.4

- *For the procurement of Non-expendable Supplies and Services:*  
The Bidder must have completed, for the last 8 years as specified in the Invitation to Bid and ITB Clause 12.1(a)(ii), a single contract that is similar to this Project, equivalent to at least twenty five (25%) of the ABC.

For this purpose, similar contracts shall refer to *Supply and Installation of Oxygen Flowmeter, Air Flowmeter and Vacuum Regulator*

*is hereby amended and shall be read as:*

- *For the procurement of Non-expendable Supplies and Services:*  
The Bidder must have completed, for the last 8 years as specified in the Invitation to Bid and ITB Clause 12.1(a)(ii), a single contract that is similar to this Project, equivalent to at **least fifty percent (50%)** of the ABC.

**Or:**

- a) Completed at least two (2) similar contracts the aggregate amount of which should be equivalent to at least fifty percent (50%) of the ABC.
- b) The largest of these similar contracts must be equivalent to at least half of the percentage of the ABC as required above.

For this purpose, similar contracts shall refer to *Supply and Installation of Oxygen Flow meter and Vacuum Regulator or Supply and Installation of Air Flow meter and Vacuum Regulator*.

*MS*  
*HS*

Republic of the Philippines  
Department of Health  
**QUIRINO MEMORIAL MEDICAL CENTER**  
Project 4, Quezon City

---

SUPPLEMENTAL / BID BULLETIN


2. Section VII. Technical Specification.

Annex A:

Technical Specification.

- Supply & Installation of Oxygen Flowmeter
    - Security Cable (safety) *is hereby amended and shall be read as:*  
Security Cable (safety)- material: Stranded steel cable, riveted connection
  - Supply & Installation of Air Flowmeter
    - Security Cable (safety) *is hereby amended and shall be read as:*  
Security Cable (safety)- material: Stranded steel cable, riveted connection
  - Supply and Installation of Vacuum Regulator with Safety Trap
    - Security Cable (safety) *is hereby amended and shall be read as:*  
Security Cable (safety)- material: Stranded steel cable, riveted connection
    - With bacteria filter per unit *is hereby added.*
    - With 40-60cm connecting tube *is hereby added.*
    - With Operational modes *is hereby added.*
    - With 1.5-2 meters connecting hose to patient *is hereby added.*
    - 0-300 to 400 mmHg negative pressure *is hereby added.*
- Other parameter not stated remains the same

For your information and guidance.

  
**ISABEL ELONA, MSc**  
Chairman, Bids and Awards Committee

